



[I-21/96/2024-W&M Section]  
GOVERNMENT OF INDIA/भारत सरकार  
MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION  
उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय  
DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग  
LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग


Krishi Bhawan/कृषि भवन, नई दिल्ली  
Dated/दिनांक:- 02.04.2024

Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by prescribed authority, is satisfied that the model described in the said report (see the figure given below) is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011, and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per ASTM E1965 & Schedule Eighth of the Legal Metrology (General) Rules, 2011 the Central Government hereby issues the certificate of approval of clinical electrical thermometer with maximum device with digital indication of accuracy 0.3 degree celsius of series "TAT-2000C" with brand name "EXERGEN" (hereafter in referred to as the said model), manufactured by M/s Exergen Corporation, 400, Pleasant Street, Watertown MA 02472, USA and imported and marketed in India without any alteration before or after sale by M/s Exergen India Private Limited, Unit No. ATM-3, G.F, BPTP Park Centra, Sec-30, Basai Road, Gurgaon-122001, Haryana and which is assigned the approval mark IND/09/24/166 (the picture of model is given below as figure 1.);

Figure-1



The said model is an electronic and infrared sensor based digital clinical infrared thermometer for intermittent determination of patient temperature with measuring range of 15.5 degree celsius to 42.0 degree celsius. It operates on 9 VDC battery direct current power supply with digital indication of temperature, scale interval or digital increment of 0.1 degree celsius and display is liquid crystal display (LCD) type.

[F.No. I-21/96/2024-W&M Section]

*Ashutosh Agarwal*  
(Ashutosh Agarwal/अशुतोष अग्रवाल)

Director (Legal Metrology) to Govt. of India/ निदेशक (विधिक माप विज्ञान) भारत सरकार

Phone/दूरभाष 01123389489

Email/ई-मेल: [dirwm-ca@nic.in](mailto:dirwm-ca@nic.in)

Online Application No. 17257