



[I-21/110/2024-W&M Section]

GOVERNMENT OF INDIA/भारत सरकार
MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION
उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय
DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग
LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग


Krishi Bhawan/कृषि भवन, नई दिल्ली

Dated/दिनांक:- 24.04.2024

Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report (see the figure given below) is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per ASTM E 1965-98 (Reapproved 2003) & Schedule Eighth of the Legal Metrology (General) Rules, 2011, the Central Government hereby issues the certificate of approval of non-contact infrared forehead clinical thermometer with digital indication of accuracy ± 0.2 degree celsius of series "TAT-2000" with brand name "EXERGEN"(hereinafter referred to as the said model), manufactured by M/s. EXERGEN CORPORATION, 400, Pleasant Street, Watertown MA 02472, USA and imported & marketed in India without making any alternation by M/s. EXERGEN INDIA PRIVATE LIMITED, Unit No ATM-3,BPTP Park Centra, Basai road, Sec-30, Gurugram Haryana 122001 and which is assigned the approval mark IND/09/24/190 (the picture of the model is given below as Figure 1);

Figure-1



The said model is a non-contact infrared forehead clinical thermometer with measuring range of 32 degree celsius to 42 degree celsius. Such instruments is intended to be used to measure the body temperature using infrared sensors. It can detect the body temperature of all ages human's forehead. The infrared thermometer can be used by consumers in household environment. It operates on 9V DC power supply with digital indication of temperature, scale interval or digital increment of 0.1 degree celsius and display is liquid crystal display (LCD) type.

[F.No.I-21/110/2024-W&M Section]

Ashutosh Agarwal

(Ashutosh Agarwal/आशुतोष अग्रवाल)
Director (Legal Metrology) to Govt. of India/निदेशक (विधिक माप विज्ञान) भारत सरकार
Phone/दूरभाष 01123389489
Email/ई-मेल: dirwm-ca@nic.in

Online application no. 17246