



[I-21/42/2025-W&M Section]
GOVERNMENT OF INDIA/भारत सरकार
MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION
उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय
DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग
LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग

Krishi Bhawan/कृषि भवन, नई दिल्ली

Dated/दिनांक- : 09.06.2025

Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per Schedule Seventh of the Legal Metrology (General) Rules, 2011, the Central Government hereby issues the certificate of approval of the model of non-automatic weighing instrument with digital indication of medium accuracy (accuracy class-III) of series "TH-WLS" and with brand name "THEMISTO" (hereinafter referred to as the said model), manufactured by M/s JIPVI ECOMMERCE Private Limited, B - 1/7 KH-17, Nanhey Park, Uttam Nagar, South West Delhi - 110059 and which is assigned the approval mark IND/09/25/ 214 (the picture of model is given below as figure 1);

The said model is a strain gauge type load cell of Make: IV Innovations Pvt Ltd, Model/Series: IV-L50, Accuracy Class: C, Max Capacity of load cell : 60kg based non-automatic weighing instrument with a maximum capacity of 50kg and minimum capacity of 1kg. The verification scale interval (e) is 50g. The liquid crystal display (LCD) display indicates the weighing result. The instrument works on 3V (Battery Operated) power supply. It is used as Electronic Luggage Type weighing instrument.

Figure 1





Back Side




Figure 2

The instrument is sealed using a tearable sticker at the back side. The contents of this scale are fully sealed and cannot be accessed without tampering the scale. For access to internal circuit & PCB, it is permanently restricted during the manufacturing process.

Further, in exercise of the power conferred by sub-rule (9) of rule 8 of the Legal Metrology (Approval of Models) Rules, 2011, the Central Government hereby declares that this certificate of approval of the said model shall also cover the weighing instruments of similar make, accuracy and performance of same series with capacity up to 60kg with number of verification scale interval (n) in the range of 500 to 10000 for 'e' value of 5g or more and with 'e' value of 1×10^k , 2×10^k or 5×10^k , where k is a positive or negative whole number or equal to zero manufactured by the same manufacturer in accordance with the same principle, design and with the same materials with which, the said approved model has been manufactured.

[F.No. I-21/42/2025-W&M Section]

Ashutosh Agarwal

Director (Legal Metrology) to Govt. of India/ निदेशक (विधिक माप विज्ञान) भारत सरकार

Phone/दूरभाष 01123389489

Email/ई-मेल: dirwm-ca@nic.in

Online Application No. 18412