



[I-21/173/2025-W&M Section]  
GOVERNMENT OF INDIA/भारत सरकार  
MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION  
उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय  
DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग  
LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग

Krishi Bhawan/कृषि भवन, नई दिल्ली

Dated/दिनांक:- 25.06.2025

Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;


Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per Schedule Eighth of the Legal Metrology (General) Rules, 2011, the Central Government hereby issues the certificate of Measuring systems for liquids other than water with digital indication of Accuracy Class 0.5 of series "ANT-23-IOT-D002" with brand name "ANT 3i TECHNOLOGIES LLP" (here in after referred to as the said model), manufactured by M/s. ANT 3i TECHNOLOGIES LLP, shop No 25, Ground Floor, ML Square, Kudalwadi, Chikhali, Vitthal Nagar, Pimpri- Chinchwad , Pune-411062 and which is assigned the approval mark IND/09/25/ 231 (the picture of the model is given below as Figure 1);

Technical Data:

Accuracy class	0.5
Maximum Flow Rate	40 lpm
Minimum Flow Rate	20 lpm
Minimum Measured Quantity(MMQ)	2 l
Resolution/Smallest Division	10 ml
Non resettable Totalizer	8 digits
Volume Display	8 digits
Amount Display	8 digits
Rate Display	5 digits
Display Type	LCD
No.of Nozzles	1
Power Supply	230 V AC / 12 V DC
Intended use	To deliver the petroleum products over IoT

Figure 1






SEAL 1-PULSER SENSOR

SEAL 2-CRANK SHAFT

SEAL 3-WHEEL FOR ADJUSTING CRANK  
SHAFT MOVEMENT



SEAL 4-MAINBOARD




Figure 2.

A typical schematic diagram of sealing provision to prevent the fraudulent practices of the model is given above as Figure 2.

[F.No. I-21/173/2025-W&M]

*Ashutosh Agarwal*

(Ashutosh Agarwal) / (आशुतोष अग्रवाल)  
Director (Legal Metrology)/निदेशक (विधिक माप विज्ञान)

Phone/दूरभाष: 01123389489

Email/ई-मेल: [dirwm-ca@nic.in](mailto:dirwm-ca@nic.in)

Online Application: 18780