



[I-21/18/2025-W&M Section]

GOVERNMENT OF INDIA/भारत सरकार
MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION
उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय
DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग
LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग

Krishi Bhawan/कृषि भवन, नई दिल्ली
Dated/दिनांक:-: 28 .01.2025

Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by prescribed authority, is satisfied that the model described in the said report (see the figure given below) is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per Schedule Sixth of the Legal Metrology (General) Rules, 2011 the Central Government hereby issues the certificate of approval of plastic tape measures of accuracy class-III of series "FG Leo 30" and with brand name "Sherla Measuring Tools"(hereinafter referred to as the said model) manufactured by M/s Sherla Measuring Tools, 132/29, First Floor, Parvati Paytha, Vithalwadi Road, Near Nilayam Bridge, Pune-411030, Maharashtra-411030 and which is assigned the approval mark IND/09/25/21 (the picture of the model is given below as Figure 1)

The said model is a plastic tape measure which is used for measurement of length where the use of rigid length measures is not convenient or practicable. The maximum length is 30m and value of graduation is 2mm and the width of the plastic tape measure is 13mm. The results of measurements are indicated by graduation lines on the plastic tape measure.

Figure 1






Figure 2

The stamp of verification shall be affixed on the metal/plastic strip provided at the beginning of the tape measure as shown in the figure above.

[F.No. I-21/18/2025-W&M Section]

Ashutosh Agarwal

(Ashutosh Agarwal/आशुतोष अग्रवाल)
Director (Legal Metrology) to Govt. of India/ निदेशक (विधिक माप विज्ञान) भारत सरकार
Phone/दूरभाष 01123389489
Email/ई-मेल: dirwm-ca@nic.in

Online application no. 21016