

## [I-21/296/2025-W&M Section] GOVERNMENT OF INDIA/भारत सरकार MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण् मंत्रालय

DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग

Krishi Bhawan/कृषि भवन, नई दिल्ली Dated/दिनांक- : 29. 10.2025.

## Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per Schedule Seventh of the Legal Metrology (General) Rules, 2011, the Central Government hereby issues the certificate of approval of the model of non-automatic weighing instrument with digital indication of series "PET" and with brand name "PRISTON" (hereinafter referred to as the said model), manufactured by M/s Priston Electro INC, 3rd Floor, Shop No. 305, Meriton Plaza, Yogi Chowk Road, Opp. Varachha Bank, Nr. Savaliya Circle, Surat, Gujarat - 395010 and which is assigned the approval mark IND/09/25/438 (the picture of model is given below as figure 1);

## **Technical Data:**

Technical Data:	
Type of Instrument	Electronic Table top type
Accuracy Class	Special Accuracy (I)
Maximum Capacity	600g
Minimum Capacity	1g
Verification Scale Interval	10mg
Power Supply	230V, 50Hz
Display Unit	LCD
Tare	Subtractive
Working Principle	Strain Gauge Type Load Cell Based
Load Cell Details	Make - LCT, Model - SS600, Class - C2, Cap - 1.2kg
Software Version	mothercard_firmware_p1.6.0.hex
Checksum	0859a6cd501628392f52fc583b674e18e41854e04ae0f1 d312f17c0d1fbce682














Figure 2

Sealing is done by passing the sealing wire through the holes provided on the body of machine. A lead seal is put at the end of the wire for stamping purpose (Sealing diagram and photograph to prevent the fraudulent practices of the model is given above as Figure 2). A dip switch has also been provided in mother board/PCB to disable access to external calibration.

Further, in exercise of the power conferred by sub-rule (9) of rule 8 of the Legal Metrology (Approval of Models) Rules, 2011, the Central Government hereby declares that this certificate of approval of the said model shall also cover the weighing instruments of similar make, accuracy and performance of same series with capacity up to 20kg with number of verification scale interval (n) 50000 or more for 'e' value of 1 mg or more with 'e' value of  $1 \text{x} 10^k$ ,  $2 \text{x} 10^k$  or  $5 \text{x} 10^k$ , where k is a positive or negative whole number or equal to zero manufactured by the same manufacturer in accordance with the same principle, design and with the same materials with which, the said approved model has been manufactured.

[F.No. <u>I-21/296/2025-W&M Section</u>]



(Ashutosh Agarwal/आशुतोष अग्रवाल) Director (Legal Metrology) to Govt. of India/ निदेशक (विधिक माप विज्ञान) भारत सरकार Phone/दूरभाष01123389489 Email/ई-मेल: dirwm-ca@nic.in

Online Application No. 23499