



[I-21/192/2025-W&M Section]
GOVERNMENT OF INDIA/भारत सरकार
MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION
उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय
DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग
LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग

Krishi Bhawan/कृषि भवन, नई दिल्ली
Dated/दिनांक:- 08 .07.2025

Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per Schedule Sixth of the Legal Metrology (General) Rules, 2011, the Central Government hereby issues the certificate of approval of the model of the plastic tape measures of accuracy class-III of series "TI-INFANTO" and with brand name "TECHNOMATE" (hereinafter referred to as the said model) manufactured by M/s Aadhunik Engineering , Tupudana, Hatia, Ranchi , Jharkhand 834003 and which is assigned the approval mark IND/09/25/271 (the picture of the model is given below as Figure 1);

The said model is a plastic tape measures which is used for measurement of length. The range is 10cm to 100cm and value of graduation is 1mm. The results of measurements are indicated by graduation lines on the tape. It may be used as infantometer for measuring length.

Figure 1






Figure 2

Sealing is done by sealing wire passed through the base and measuring scale if the device and lead stud is affixed for seal. Sealing is done at bottom side of the instrument. A typical schematic diagram of sealing provision to prevent the fraudulent practices of the model is given above as Figure 2.

[F.No. I-21/192/2025-W&M Section]



(Ashutosh Agarwal/आशुतोष अग्रवाल)
Director (Legal Metrology) to Govt. of India/निदेशक (विधिक माप विज्ञान) भारत सरकार
Phone/दूरभाष 01123389489
Email/ई-मेल: dirwm-ca@nic.in

Online Application No. 23344