



[I-21/157/2025-W&M Section]

GOVERNMENT OF INDIA/भारत सरकार
MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION

उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय
DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग
LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग

Krishi Bhawan/कृषि भवन, नई दिल्ली

Dated/दिनांक:- 10.06.2025

Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by prescribed authority, is satisfied that the model described in the said report (see the figure given below) is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;


Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per OIML R 49, the Central Government hereby issues the certificate of approval of the model of water meter (Ultrasonic) of accuracy class 2 (hereinafter referred to as the said model) of series "NX-UWM" and with brand name "IDEAL" manufactured by M/s. Nixies Meters Private Limited, Plot No 32, EPIP, Industrial Layoutm Ganjimutt, Mangalore, Dakshina Kannada Karnataka-574144, and which is assigned the approval mark IND/09/25/ 215 (the picture of the model is given below as Figure 1);

The said model is a water meter (ultrasonic) which is based on principle of ultrasonic sensing.

Technical Data:

Accuracy Class	2
DN Size	15 mm
Flow Rates	$Q_1 = 0.006 \text{ m}^3/\text{h}$ $Q_2 = 0.010 \text{ m}^3/\text{h}$ $Q_3 = 2.5 \text{ m}^3/\text{h}$ $Q_4 = 3.125 \text{ m}^3/\text{h}$
Type of Display	LCD
No. of digits	9
Power Supply	3.6 V DC

Figure-1





Place of sealing

Figure 2

In the electronic unit model, a disposable cover is used, the opening of which is impossible without damages. Protection against unauthorized access is provided by the design of the meters. A typical schematic diagram of sealing provision to prevent the fraudulent practices of the model is given above as Figure 2.

[F.No. I-21/157/2025-W&M Section]

Ashutosh Agarwal

Director (Legal Metrology) to Govt. of India/ (Ashutosh Agarwal/आशुतोष अग्रवाल)
निदेशक (विधिक माप विज्ञान) भारत सरकार

Phone/दूरभाष 01123389489

Email/ई-मेल: dirwm-ca@nic.in

Online Application No. 22183