



[I-21/4/2026-W&M Section]

GOVERNMENT OF INDIA/भारत सरकार

MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION

उपभोक्ता मामले, खाद्य एवं सावजनिक वितरण मंत्रालय

DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग

LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग

Krishi Bhawan/कृषि भवन, नई दिल्ली

Dated/दिनांक- : 09.01.2026

Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per Schedule Seventh of the Legal Metrology (General) Rules, 2011, the Central Government hereby issues the certificate of approval of the model of non-automatic weighing instrument with digital indication of series "MMTT" and with brand name "MAHESHWARI SCALES" (hereinafter referred to as the said model), manufactured by M/s Mangal Murti Weighing Solutions, 464/133B, Madhowadi Jatawpura, Near Bank of Baroda, Narkulganj Branch, Bareilly - 243005. Uttar Pradesh and which is assigned the approval mark IND/09/26/ 23 (the picture of model is given below as figure 1);

Technical Data:

Type of Instrument	Electronic Table top Type
Accuracy Class	High Accuracy (II)
Maximum Capacity	30kg
Minimum Capacity	100g
Verification Scale Interval	2g
Power Supply	230V 50Hz
Display Unit	Seven Segment LED
Tare	Subtractive
Working Principle	Strain Gauge type load cell based
Load Cell	Make-LOADMASTER,CAPACITY-50kg,Class-C3, Model- BT 50 LM
Software Version	mothercard firmware v1.5.0.hex
Checksum	e24a8d3c12b7f5a89e1b2f7c0d95e6f3b7c3d2f9a1e0b5c 4d6a7b8c9d0e1f2a3





Figure 1



Figure 2



METAL WIRE INSERTED
AND SEALED WITH LEAD
MATERIAL.



Sealing is done by passing the sealing wire through the holes provided on the body of machine. A lead seal is put at the end of the wire for stamping purpose (Sealing diagram and photograph to prevent the fraudulent practices of the model is given above as Figure 2). A dip switch has also been provided in mother board/PCB to disable access to external calibration.

Further, in exercise of the power conferred by sub-rule (9) of rule 8 of the Legal Metrology (Approval of Models) Rules, 2011, the Central Government hereby declares that this certificate of approval of the said model shall also cover the weighing instruments of similar make, accuracy and performance of same series with capacity up to 50kg with number of verification scale interval (n) in the range of 5000 to 60000 for 'e' value of 100mg or more and with number of verification scale interval (n) in the range of 100 to 60000 for 'e' value of 1mg to 50mg with 'e' value of 1×10^k , 2×10^k or 5×10^k , where k is a positive or negative whole number or equal to zero manufactured by the same manufacturer in accordance with the same principle, design and with the same materials with which, the said approved model has been manufactured.

[F.No. I-21/4/2026-W&M Section]



(Ashutosh Agarwal/आशुतोष अग्रवाल)
Director (Legal Metrology) to Govt. of India/ निदेशक (विधिक माप विज्ञान) भारत सरकार
Phone/टूरभाष 01123389489
Email/ई-मेल: dirwm-ca@nic.in

Online Application No. 23662