



Online Application No 24004

[I-21/219/2026-W&M Section]

GOVERNMENT OF INDIA/भारत सरकार
MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION
उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय
DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग
LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग

Krishi Bhawan/कृषि भवन, नई दिल्ली

Dated/दिनांक- : 10. 07.2026.

Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per Schedule Seventh of the Legal Metrology (General) Rules, 2011, the Central Government hereby issues the certificate of approval of the model of non-automatic weighing instrument with digital indication of series "IND 400" and with brand name "METTLER TOLEDO" (hereinafter referred to as the said model), manufactured by M/s METTLER- TOLEDO INDIA Pvt, Ltd., Amar Hill, Saki Vihar Road, Powai, MUMBAI-400072 and which is assigned the approval mark IND/09/26/338 (the picture of model is given below as figure 1);

Technical Data:

Type of Instrument	Electronic Platform Weighing Scale
Accuracy Class	High Accuracy Class II
Maximum Capacity	300kg
Minimum Capacity	1000g
Verification Scale Interval	20g
Scale Interval 'd'	2g
Power Supply	230 V, 50Hz AC / 12 V DC
Display Unit	TFT colour Graphic Display
Tare	Subtractive
Working Principle	Strain Gauge Load Cell Based
Load Cell	Make -Mettler Toledo, 0745A, Accuracy Class: C3, Max Cap : 220kg , 04 nos.
Software Version	2.xx.yyyy
Checksum Details	4261bebd2dfc08c3f737c99e756620f1150db15ce3b1df6 df7d0e2efc77f213062a8077c3c119864a832e1aff406db a960005df0c3e03b3a08de205dbe24c22



Figure






Figure 2

Sealing is done by passing the sealing wire through the holes provided on the body of machine. A lead seal is put at the end of the wire for stamping purpose (a typical schematic diagram of sealing provision to prevent the fraudulent practices of the model is given above as Figure 2). A dip switch has also been provided in mother board/PCB to disable access to external calibration.

Further, in exercise of the power conferred by sub-rule (9) of rule 8 of the Legal Metrology (Approval of Models) Rules, 2011, the Central Government hereby declares that this certificate of approval of the said model shall also cover the weighing instruments of similar make, accuracy and performance of same series with capacity above 50kg and up to 5000kg with number of verification scale interval (n) in the range of 5000 to 60000 for 'e' value of 100mg or more with 'e' value of 1×10^k , 2×10^k or 5×10^k , where k is a positive or negative whole number or equal to zero manufactured by the same manufacturer in accordance with the same principle, design and with the same materials with which, the said approved model has been manufactured.

[F.No. I-21/219/2026-W&M Section]



(Ashutosh Agarwal/आशुतोष अग्रवाल)
Director (Legal Metrology) to Govt. of India/ निदेशक (विधिक माप विज्ञान) भारत सरकार
Phone/दूरभाष-01123389489
Email/ई-मेल: dirwm-ca@nic.in