



Online Application No 24028

[I-21/209/2026-W&M Section]

GOVERNMENT OF INDIA/भारत सरकार
MINISTRY OF CONSUMER AFFAIRS, FOOD AND PUBLIC DISTRIBUTION

उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय
DEPARTMENT OF CONSUMER AFFAIRS/ उपभोक्ता मामले विभाग

LEGAL METROLOGY DIVISION/ विधिक मापविज्ञान प्रभाग

Krishi Bhawan/कृषि भवन, नई दिल्ली

Dated/दिनांक- : 23.07.2026

Certificate of Approval of Model/ मॉडल का अनुमोदन प्रमाणपत्र

Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, as per Schedule Seventh of the Legal Metrology (General) Rules, 2011, the Central Government hereby issues the certificate of approval of the model of non-automatic weighing instrument with digital indication of series “KaradaScan 375 Plus” and with brand name “OMRON” (hereinafter referred to as the said model), manufactured by M/s Omron Healthcare Co. Ltd., 53 Kunotsubo, Teradocho, Muko, Kyoto, 617-002, Japan and imported and marketed before and after sale without any alteration in India by M/s Omron Healthcare India Private Limited, 6th Floor, B-Block, Sewa Tower, Plot No: 19, Sector-18, Udyog Vihar, Maruti Industrial Complex, Haryana, Gurugram-122008 which is assigned the approval mark IND/09/26/ 360 (the picture of model is given below as figure 1);

Technical Data:

Type of Instrument	Person Weighing Type
Accuracy Class	Medium Accuracy (III)
Maximum Capacity	100 kg/135kg
Minimum Capacity	2kg
Verification Scale Interval	100g/200g
Power Supply	6V DC
Display Unit	LCD
Tare	Subtractive
Working Principle	Strain Gauge type load cell based
Load Cell	Make: KRELL PRECISION(YANGZHOU) CO, LTD, Model::FS-018, Accuracy Class: C-3, Max Capacity: 240kg
Software Version	For Main CPU (BH67F2495) 004 (Software Version) For Sub CPU (BH66F2663) 005(Software Version)
Checksum Details	For Main CPU (BH67F2495) 0xF2E6 For Sub CPU (BH66F2663) 0x1135



Figure 1






Figure 2

Sealing is done by passing the sealing wire through the holes provided on the body of machine. A lead seal is put at the end of the wire for stamping purpose (a typical schematic diagram of sealing provision to prevent the fraudulent practices of the model is given above as Figure 2). A dip switch has also been provided in mother board/PCB to disable access to external calibration.

[F.No. I-21/209/2026-W&M Section]



(Ashutosh Agarwal/आशुतोष अग्रवाल)
Director (Legal Metrology) to Govt. of India/ निदेशक (विधिक माप विज्ञान) भारत सरकार
Phone/दूरभाष-01123389489
Email/ई-मेल: dirwm-ca@nic.in